

CERTIFICATE OF ANALYSIS



product: CHROMagar MRSA

Cat. No.: 201402 Lot No.: 20101104 Expiry date: 2010-12-30 Type of medium: selective-differential

Formula / Liter:		Supplements / Liter:	
Peptone and yeast extract	40,0 g	CHROMagar MRSA Supplement	0,9 ml
Salts	25,0 g		
Chromogenic mix	2,5 g		
Agar	15,0 g		

Characteristic of medium:	Spec. Values:	Batch Values:
color	yellow beige	confirmed
clarity / homogeneity	clear	confirmed
consistency	gel	confirmed
pH	6,9 ± 0,2	6,9

Sterility control:	Spec. Values:	Batch Values:
incubation time: 3 - 5 days	no growth	confirmed
incubation temp.: 37 and 25 ± 2°C		

Incubation time and temperature:	
time:	24h
temperature:	35 ± 2°C

Microorganism:	colony appearance:	recovery (PR):	growth:
<i>Staphylococcus aureus</i> ATCC 43300	rose to mauve	0,89	good
<i>Staphylococcus aureus</i> ATCC 25923	—	—	no growth

Reference medium - Tryptic Soy Agar

Quality control was executed according to ISO 11133-2:2003 standard.

Test performed and confirmed by: Gliniecka Anna

Graso, 83-200 Starogard Gd, Krag 4a

www.podloza.pl

